

Certificate of Analysis

Product Description: HBc IgM Positive Greater Than >4 S/CO Plasma
 Material Number: 0305-0017, 0305-0018, 0305-0019, 0305-0020
 Batch: 10538067
 Donor ID: BD459693
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 23 SEP 2020
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
09-Apr-21	HBc IgM	Abbott Architect	7.120	S/CO	1:8 dilution
05-Mar-21	HBc IgM	DiaSorin Liaison XL	>15	INDEX	NA
This product was tested for Antibodies for HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
12-Apr-21	HBeAg	DiaSorin EIA Plus	63.492	S/CO	1:8 dilution
05-Mar-21	HBeAg	DiaSorin Liaison XL	>1,200	INDEX	NA
08-Mar-21	HBsAg	Bio-Rad EIA	38.095	S/CO	OFF SCALE
23-Mar-21	HBV DNA	Roche cobas 6800	45,300,000.000	IU/ML	NA

NA = Not Available

Approval:

Lynne Labonte

16 JUN 2021

Prepared By

Date

Danni Parker

21 JUN 2021

QA Verified By

Date