

Certificate of Analysis

Product Description: HBc IgM Positive Greater Than 4 S/CO Plasma
 Material Number: 0305-0017, 0305-0018, 0305-0019, 0305-0020
 Batch: 9219955
 Donor ID: BD200251
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 03 NOV 2007
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
24 Feb 2022	HBc IgM	Abbott Architect	6.970	S/CO	NA
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV and found to be negative.					
Caution: HBsAg Positive					
Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
17 Feb 2022	HBeAg	DiaSorin Liaison XL	0.498	INDEX	≥0.9 = POS
21 May 2020	HBsAg	Abbott Architect	4.500	S/CO	NA
24 Feb 2022	HBsAg	Abbott Architect	2.050	S/CO	NA

NA = Not Available

Approval:

Syrene Labonte

26 MAR 2024

Prepared By

Date

Paula J. Kuman

27 MAR 2024

QA Verified By

Date