

Certificate of Analysis

Product Description: HBc IgM Positive Greater Than 4 S/CO Plasma
 Material Number: 0305-0017, 0305-0018, 0305-0019, 0305-0020
 Batch: 9222139
 Donor ID: BD200258
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 08 DEC 2007
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.
LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
15 Feb 2008	HBc IgM	Abbott EIA	7.270	S/CO	NA
28 Nov 2011	HBc IgM	DiaSorin EIA Plus	11.026	S/CO	NA

This product was tested for Antibodies to HIV 1/2, Antibodies to HCV and found to be negative.

Caution: HBsAg Positive

NA = Not Available

Approval:

Lynne Labonte

26 MAR 2024

Prepared By

Date

Gaule J. Keman

27 MAR 2024

QA Verified By

Date