

Certificate of Analysis

Product Description: EBV VCA IgM Positive Plasma
 Material Number: 0345-0001, 0345-0003, 0345-0004, 0345-0005
 Batch: 10624943
 Donor ID: BD500080
 Gender: NA
 Age: NA
 Country of Origin: United States
 Collection Date: 10 JUN 2022
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
27 Jun 2022	EBV VCA IgM	Zeus EIA	7.048	S/CO	NA
03 Jun 2024	EBV VCA IgM	DiaSorin Liaison XL	32.4	U/mL	>44 U/mL is POS
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.					

NA = Not Available

Approval:

Lynne Rabente

10 JUN 2024

Prepared By

Date

Gaula J. Kuman

11 JUN 2024

QA Verified By

Date