

Certificate of Analysis

Product Description: AccuPlex™ H5N1 Influenza Reference Material Kit

Material Number: 0505-0422

Lot Number: 10742250

Date of Manufacture: 21FEB2025

Expiration Date: 17FEB2027

Storage Temperature: 2 – 8°C


Viral copies/mL as determined by digital PCR assays run on the Bio-Rad QX200 Droplet Digital PCR System (for Research Use Only):

QC Testing:

Kit Component	Digital PCR Target Sequence(s)	Results (cp/mL)
Positive Reference Material	Influenza H5N1	5,884
Negative Reference Material	Human RNaseP	6,533

Safety Precautions: Preparation contains recombinant Sindbis virus. Use the Center for Disease Control recommended universal precautions.

Approval:


Prepared By

25FEB2025

Date


QA Verified By

27Feb2025

Date