

Certificate of Analysis

Product Description:	Seraseq® ctDNA Mutation Mix v4 AF0.5%		
Material Number:	0710-3099	Batch Number:	10687646
Material Description:	A ctDNA-like mixture of human genomic DNA from the reference cell line, GM24385, and synthetic DNA constructs		
Volume:	25 µL		
Date of Manufacture:	24 MAY 2024	Expiration Date:	12 APR 2028
Storage	-20°C		
Nominal Concentration:	10 ng/µL		
Concentration (Qubit dsDNA BR Assay):	12.4 ng/µL		
Average Fragment Size (Agilent Bioanalyzer High Sensitivity DNA Analysis):	207 bp		
Acceptance Criteria for Fragment Size:	150-220 bp		
Digital PCR Method:	Bio-Rad QX200™ Droplet Digital™ PCR System		
NGS Method:	Illumina TruSight™ Oncology 500 ctDNA sequenced on NovaSeq™		

Certificate of Analysis

Product Description: Seraseq® ctDNA Mutation Mix v4 AF0.5%

Material Number: 0710-3099

Batch Number:

10687646

SNVs and Indels

Gene	Nucleotide Change	Protein Change	dPCR VAF%	NGS VAF%
AKT1	c.49G>A	p.E17K	0.513	0.51
AR	c.2623C>T	p.H875Y	1.163*	0.00*
ATM	c.1058_1059del	p.C353Sfs*5	0.553	0.47
BRAF	c.1799T>A	p.V600E	0.480	0.69
BRCA1	c.1961del	p.K654Sfs*47	0.437	0.42
BRCA2	c.7934del	p.R2645Nfs*3	0.360	0.33
CDKN2A	c.9_32dup	p.A4_P11dup	0.370	NA
CHEK1	c.676del	p.T226Hfs*14	0.430	NA
CHEK2	c.1029_1030delinsGT	p.K373*	0.320	NA
EGFR	c.2369C>T	p.T790M	0.523	0.47
EGFR	c.2303G>T	p.S768I	0.340	0.36
EGFR	c.2389T>A	p.C797S	0.523	0.59
EGFR	c.2573T>G	p.L858R	0.447	0.41
EGFR	c.2310_2311insGGT	p.D770_N771insG	0.340	0.36
EGFR	c.2235_2249del	p.E746_A750del	0.417	0.39
ERBB2	c.2313_2324dup	p.Y772_A775dup	0.447	0.73
ESR1	c.1613A>G	p.D538G	0.387	0.61
FGFR3	c.746C>G	p.S249C	0.420	0.27
HRAS	c.37G>C	p.G13R	0.477	0.42
HRAS	c.182A>G	p.Q61R	0.477	0.24
IDH1	c.394C>T	p.R132C	0.397	0.35
IDH2	c.419G>A	p.R140Q	0.423	0.28
IDH2	c.515G>A	p.R172K	0.460	0.35
KIT	c.2361+67_2361+72delTTTTTT	N/A	0.390	NA
KIT	c.2447A>T	p.D816V	0.427	0.71
KRAS	c.183A>C	p.Q61H	0.367	0.35
KRAS	c.34G>T	p.G12C	0.397	0.37
KRAS	c.35G>A	p.G12D	0.390	0.70
MAP2K1	c.370C>T	p.P124S	0.440	0.22
MAP4K3	c.246-2475_246-2470delTTTTTT	N/A	0.597	NA
MAP4K3	c.998-35_998-30delAAAAAA	N/A	0.460	NA
MET	c.3082+1del	N/A	0.430	0.38
MLH1	c.232_243delinsATGTAAGG	p.E78_T81delinM*	0.590	NA
MSH2	c.1662-12_1677del	N/A	0.560	0.23
MSH2	c.942+20_942+29delAAAAAAAAAA	N/A	0.507	NA

*The AR variant is present on the X chromosome and the genomic background of this product is male, thus VAF will be ~2-fold higher than other variants.

NA = not assayed; not covered by NGS test panel

Certificate of Analysis

Product Description: Seraseq® ctDNA Mutation Mix v4 AF0.5%

Material Number: 0710-3099

Batch Number:

10687646

SNVs and Indels, continued

Gene	Nucleotide Change	Protein Change	dPCR VAF%	NGS VAF%
MSH6	c.2056_2060delinsCTTCTACCTCAAAAA	p.G686Lfs*10	0.420	NA
MTOR	c.6644C>A	p.S2215Y	0.500	0.55
NF1	c.3738_3747del	p.F1247Gfs*16	0.390	NA
NRAS	c.182A>G	p.Q61R	0.553	0.47
NTRK1	c.1783G>A	p.G595R	0.550	0.35
NTRK2	c.1915G>A	p.G639R	0.470	0.38
NTRK3	c.1867G>A	p.G623R	0.527	0.76
PALB2	c.839del	p.N280Tfs*8	0.445	0.65
PDGFRA	c.2525A>T	p.D842V	0.463	0.45
PIK3CA	c.3140A>G	p.H1047R	0.473	0.46
PIK3CA	c.3203dup	p.N1068Kfs*5	0.473	0.44
PIK3CA	c.1633G>A	p.E545K	0.400	0.13
PIK3R1	c.1727_1729del	p.T576del	0.390	0.48
PMS2	c.861_864del	p.R287Sfs*19	0.560	0.33
PTCH1	c.2307_2308delinsTT	p.R770*	0.380	0.56
PTEN	c.800del	p.K267Rfs*9	0.483	0.30
PTEN	c.741dup	p.P248Tfs*5	0.483	0.21
RAD51C	c.338dup	p.G114Wfs*25	0.428	NA
RAD51C	c.242C>A	p.S81*	0.428	0.20
RAD51D	c.392dup	p.N131Kfs*23	0.420	0.56
RAD51D	c.271A>T	p.K91*	0.420	0.32
RAF1	c.770C>T	p.S257L	0.390	0.48
RB1	c.751C>T	p.R251*	0.393	0.33
RET	c.2753T>C	p.M918T	0.507	0.51
SLC7A8	c.-231_-224delTTTTTTTT	N/A	0.430	NA
SMARCB1	c.118C>T	p.R40*	0.500	0.55
STK11	c.734+1G>T	N/A	0.480	0.42
TERT	c.-124C>T	N/A	0.400	NA
TERT	c.-146C>T	N/A	0.400	NA
TP53	c.723del	p.C242Afs*5	0.477	0.34
TP53	c.743G>A	p.R248Q	0.477	0.34
TP53	c.818G>A	p.R273H	0.477	0.31
TSC1	c.1263+1G>T	N/A	0.510	0.44
TSC2	c.2640-1G>A	N/A	0.603	0.33
VHL	c.481C>T	p.R161*	0.520	0.38
ZNF2	c.*1525_*1530delTTTTTT	N/A	0.400	NA

NA = not assayed; not covered by NGS test panel

Certificate of Analysis

Product Description: Seraseq® ctDNA Mutation Mix v4 AF0.5%
Material Number: 0710-3099 **Batch Number:** 10687646

Structural Variants

Genes	Junction	dPCR AF%	NGS AF%
CD74::NRG1	Intron 6::Intron 5	0.590	0.18
CD74::ROS1	Intron 6::Intron 34	0.473	0.32
COL1A1::PDGFB	Intron 25::Intron 1	0.320	NA
EML4::ALK	Intron 13::Intron 19	0.470	0.33
ETV6::NTRK3	Intron 5::Intron 14	0.450	NA
FGFR2::BICC1	Intron 17::Intron 2	0.387	0.16
FGFR3::TACC3	Exon 18::Intron 7	0.470	0.69
NCOA4::RET	Intron 7::Intron 11	0.353	0.30
PML::NTRK2	Intron 2::Intron 12	0.520	NA
TPM3::NTRK1	Intron 7::Intron 9	0.340	NA

NA = not assayed

Copy Number Variants

Gene	dPCR Total Copies	NGS Total Copies
AKT2	2.49	2.42
CCND1	2.60	2.88
CCNE1	2.66	3.39
CDK4	2.58	2.66
ERBB2	2.51	2.41
FGF19	2.60	2.77
FGF3	2.60	2.35
FGF4	2.60	2.40
FGFR1	2.53	2.44
MET	2.65	3.26
MYC	3.25	7.34

Approval:

Prepared By *Tyler* Date 06/27/2024
 QA Verified By *ES* Date 28 JUN 2024