

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® Trisomy 18 Male - Matched Reference Material
Material Number: 0720-0171 Batch Number: 10693471
Material Description: A ccfDNA-like mixture of human genomic DNA (matched maternal and fetal) in a commutable matrix (simulated plasma).
Concentration (Qubit dsDNA BR Assay): 31.6 ng/mL
Volume: 1.0 mL
Date of Manufacture: 08 APR 2024 Expiration Date: 17 MAR 2026
Fetal Fraction by NGS (Sage Prenatal Screen): 7.5%

Trisomy Call for Chromosome 18 by NGS (Sage Prenatal Screen): Pass
Call for fetal gender by NGS (Sage Prenatal Screen): XY
Storage: 2-8°C

Approval:

A handwritten signature in blue ink, appearing to be "M. J.", is written over a horizontal line.

Prepared By

23 APR 2024

Date