

# Technical Product Report

*For Research Use Only; Not for use in Diagnostic Procedures*

Product Description: Seraseq® Turner Syndrome (XO) Reference Material

Material Number: 0720-0952                      Batch Number: 10732306

Material Description: A ccfDNA-like mixture of human genomic DNA (matched maternal and fetal) in a commutable matrix (simulated plasma).

Concentration  
(Qubit dsDNA BR Assay): 31.6 ng/mL

Fill Volume: 1.0 mL

Date of Manufacture: 17 DEC 2024                      Expiration Date: 16 OCT 2028


Method used for NIPT: Chromapanel

Fetal Fraction: 12.0%

Call for XO: Pass

Storage: 2-8 °C

**Approval:**

  
Prepared By

20-DEC-2024  
Date