

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq® Triple X Syndrome Reference Material

Material Number: 0720-1069

Batch Number: 10743912

Material Description: A cfDNA-like mixture of human genomic DNA (matched maternal and fetal) in a commutable matrix (simulated plasma)

Date of Manufacture: 03 MAR 2025

Expiration Date: 25 JUL 2027

Fill Volume: 1.0 mL

Storage: 2-8 °C

Concentration
(Invitrogen™ Qubit™
dsDNA BR Assay): 31.9 ng/mL

Method Used for
NIPT: Illumina VeriSeq™ NIPT Solution v2

Fetal Fraction: 21.0 %

Call for Triple X: Pass

Approval:

Prepared By



Date

13-MAR-2025